

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090322\
 Data File : PS020437.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2022 03:35
 Operator : AJ\MA
 Sample : PB147429BS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB147429BS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/06/2022
 Supervised By : Ankita Jodhani 09/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 04:32:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090222.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 03 01:09:33 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.174	6.757	1898.2E6	583.2E6	538.536m	542.762
Target Compounds							
1) T	Dalapon	2.997	2.512	2389.9E6	672.3E6	471.869	470.522
2) T	3,5-DICHL...	6.393	5.943	2348.9E6	853.5E6	492.599	506.744
3) T	4-Nitroph...	7.010	6.381	793.5E6	322.4E6	511.175	512.379
5) T	DICAMBA	7.362	6.918	7094.9E6	2466.9E6	504.333	515.475
6) T	MCP	7.531	7.013	520.3E6	184.2E6	47.583	50.100
7) T	MCPA	7.679	7.224	738.8E6	269.0E6	47.912	49.603
8) T	DICHLORPROP	8.054	7.537	1781.7E6	633.1E6	484.110	506.643
9) T	2,4-D	8.286	7.830	2327.1E6	783.0E6	496.028	515.638
10) T	Pentachlo...	8.588	8.251	25252.5E6	9158.9E6	531.343	538.453
11) T	2,4,5-TP ...	9.138	8.612	10649.3E6	3805.8E6	501.890	525.695
12) T	2,4,5-T	9.424	8.990	11417.4E6	3632.3E6	492.338	536.489
13) T	2,4-DB	9.978	9.494	2218.8E6	468.2E6	509.562	517.094
14) T	DINOSEB	11.123	9.816	11013.3E6	2673.3E6	478.145	509.023
15) T	Picloram	10.955	10.798	18773.3E6	4590.3E6	453.364	446.966
16) T	DCPA	11.393	10.743	16933.3E6	4780.6E6	527.039	536.729

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090322\
Data File : PS020437.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2022 03:35
Operator : AJ\MA
Sample : PB147429BS
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB147429BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/06/2022
Supervised By : Ankita Jodhani 09/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 04:32:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090222.M
Quant Title : 8080.M
QLast Update : Sat Sep 03 01:09:33 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

