

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
 Data File : PS027581.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 13:26
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 15:09:57 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 03 15:04:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.149	7.731	607.9E6	455.9E6	233.924	232.943
Target Compounds							
1) T	Dalapon	2.589	2.689	587.1E6	784.5E6	190.658	221.387
2) T	3,5-DICHL...	6.334	6.687	881.4E6	604.4E6	214.620	210.612
3) T	4-Nitroph...	6.947	7.260	314.5E6	262.9E6	207.732	202.882
5) T	DICAMBA	7.329	7.928	2304.5E6	1868.0E6	208.513	207.895
6) T	MCP	7.504	8.026	127.9E6	111.3E6	17.456	18.952
7) T	MCPA	7.649	8.268	200.6E6	179.2E6	19.945	20.746
8) T	DICHLORPROP	8.020	8.639	622.2E6	476.5E6	225.087	215.471
9) T	2,4-D	8.244	8.968	629.7E6	532.7E6	217.213	215.791
10) T	Pentachlo...	8.533	9.493	8934.1E6	7365.7E6	214.740	216.179
11) T	2,4,5-TP ...	9.102	9.867	3401.7E6	2874.1E6	215.173	213.411
12) T	2,4,5-T	9.389	10.285	3263.4E6	2819.7E6	212.907	212.601
13) T	2,4-DB	9.956	10.849	460.0E6	342.4E6	213.415	208.880
14) T	DINOSEB	11.138	11.227	2341.1E6	1693.0E6	208.306	203.470
15) T	Picloram	10.957	12.317	4045.9E6	3697.2E6	206.201	199.589
16) T	DCPA	11.437	12.266	4077.0E6	3138.9E6	215.947	214.229

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
Data File : PS027581.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 13:26
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 15:09:57 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
Quant Title : 8080.M
QLast Update : Tue Sep 03 15:04:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

