

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
 Data File : PS027593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 18:17
 Operator : AR\AJ
 Sample : PB163103BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB163103BS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/04/2024
 Supervised By :Ankita Jodhani 09/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 08:30:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 03 15:23:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.147	7.729	1296.3E6	981.3E6	511.191	510.478
Target Compounds							
1) T	Dalapon	2.585	2.687	1415.7E6	1634.2E6	460.612	470.979
2) T	3,5-DICHL...	6.332	6.686	1910.3E6	1341.6E6	475.770	474.243
3) T	4-Nitroph...	6.944	7.259	699.8E6	598.2E6	469.489	464.332
5) T	DICAMBA	7.328	7.927	5189.3E6	4262.5E6	477.175	479.914
6) T	MCP	7.505	8.027	350.6E6	278.0E6	47.074	46.911
7) T	MCPA	7.650	8.270	467.7E6	394.5E6	46.622	45.977
8) T	DICHLORPROP	8.017	8.638	1281.4E6	1034.4E6	475.159m	476.263
9) T	2,4-D	8.243	8.966	1365.3E6	1162.2E6	481.220	479.661
10) T	Pentachlo...	8.532	9.492	20022.0E6	16477.7E6	504.251	498.737
11) T	2,4,5-TP ...	9.101	9.866	7582.6E6	6465.4E6	490.292	489.819
12) T	2,4,5-T	9.388	10.284	7344.4E6	6387.0E6	488.805	490.747
13) T	2,4-DB	9.954	10.847	1020.6E6	779.3E6	479.589	479.577
14) T	DINOSEB	11.137	11.226	5385.9E6	3994.4E6	488.146	487.058
15) T	Picloram	10.954	12.315	9219.9E6	8826.2E6	473.890	478.685
16) T	DCPA	11.436	12.265	9219.0E6	7204.2E6	499.090	501.590

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
Data File : PS027593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 18:17
Operator : AR\AJ
Sample : PB163103BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB163103BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/04/2024
Supervised By : Ankita Jodhani 09/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 08:30:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
Quant Title : 8080.M
QLast Update : Tue Sep 03 15:23:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

