

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
 Data File : PS027598.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 20:18
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/04/2024
 Supervised By : Ankita Jodhani 09/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 08:32:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 03 15:23:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.148	7.730	1805.3E6	1377.5E6	711.935	716.581
Target Compounds							
1) T	Dalapon	2.587	2.688	2036.0E6	2188.3E6	662.428	630.676
2) T	3,5-DICHL...	6.332	6.687	2598.7E6	1898.6E6	647.209m	671.151
3) T	4-Nitroph...	6.945	7.259	962.2E6	843.2E6	645.485	654.479
5) T	DICAMBA	7.328	7.928	7406.7E6	6098.3E6	681.077	686.602
6) T	MCPD	7.507	8.029	522.6E6	409.8E6	70.170	69.151
7) T	MCPA	7.651	8.272	673.2E6	567.7E6	67.109	66.150
8) T	DICHLORPROP	8.018	8.638	1785.6E6	1479.7E6	662.142	681.288
9) T	2,4-D	8.243	8.967	1900.2E6	1643.6E6	669.747	678.366
10) T	Pentachlo...	8.531	9.493	27918.3E6	23214.2E6	703.119	702.631
11) T	2,4,5-TP ...	9.101	9.867	10567.5E6	9168.4E6	683.299	694.594
12) T	2,4,5-T	9.387	10.284	10317.2E6	9038.0E6	686.659	694.438
13) T	2,4-DB	9.953	10.848	1431.8E6	1118.3E6	672.827	688.170
14) T	DINOSEB	11.136	11.226	7630.7E6	5760.8E6	691.604	702.446
15) T	Picloram	10.953	12.316	13149.8E6	12740.3E6	675.881	690.968
16) T	DCPA	11.434	12.265	12870.9E6	10091.5E6	696.795	702.615

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
Data File : PS027598.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 20:18
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/04/2024
Supervised By : Ankita Jodhani 09/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 08:32:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
Quant Title : 8080.M
QLast Update : Tue Sep 03 15:23:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

