

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
 Data File : PS027599.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2024 20:42
 Operator : AR\AJ
 Sample : LCS-1
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LCS-1

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/04/2024
 Supervised By :Ankita Jodhani 09/04/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 08:32:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 03 15:23:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.146	7.730	925.2E6	661.9E6	364.850m	344.301
Target Compounds							
1) T	Dalapon	2.578	2.693	582.2E6	970.6E6	189.415m	279.733m#
2) T	3,5-DICHL...	6.332	6.687	1182.5E6	859.3E6	294.511m	303.771
3) T	4-Nitroph...	6.944	7.259	429.6E6	362.0E6	288.198m	280.999m
5) T	DICAMBA	7.328	7.928	3060.4E6	2654.0E6	281.415	298.806
6) T	MCP	7.503	8.027	186.0E6	158.6E6	24.968	26.763
7) T	MCPA	7.648	8.269	281.7E6	289.5E6	28.085	33.739
8) T	DICHLORPROP	8.017	8.638	851.1E6	675.4E6	315.591	310.987
9) T	2,4-D	8.242	8.967	1169.3E6	836.0E6	412.124	345.051
10) T	Pentachlo...	8.530	9.493	11530.9E6	9552.1E6	290.404	289.116
11) T	2,4,5-TP ...	9.100	9.866	4734.1E6	4140.8E6	306.109	313.708
12) T	2,4,5-T	9.386	10.284	4361.1E6	3939.0E6	290.249	302.652
13) T	2,4-DB	9.953	10.848	559.2E6	489.9E6	262.759m	301.455
14) T	DINOSEB	11.135	11.227	489.6E6	611.6E6	44.376m	74.579 #
15) T	Picloram	10.952	12.315	5559.7E6	5273.7E6	285.758m	286.019
16) T	DCPA	11.431	12.265	5490.8E6	5101.7E6	297.255	355.203

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090324\
Data File : PS027599.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2024 20:42
Operator : AR\AJ
Sample : LCS-1
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
LCS-1

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/04/2024
Supervised By : Ankita Jodhani 09/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 08:32:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090324.M
Quant Title : 8080.M
QLast Update : Tue Sep 03 15:23:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

