

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090325\
 Data File : PS031620.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2025 22:45
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/04/2025
 Supervised By :mohammad ahmed 09/05/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 02:28:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080725.M
 Quant Title : 8080.M
 QLast Update : Fri Aug 08 03:17:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.285	7.742	2677.2E6	642.5E6	843.484	760.019
Target Compounds							
1) T	Dalapon	2.673	2.692	3532.0E6	1804.5E6	672.901	678.131
2) T	3,5-DICHL...	6.453	6.695	3713.9E6	1046.9E6	787.115	774.357
3) T	4-Nitroph...	7.094	7.293	830.7E6	834.8E6	805.171	664.207
5) T	DICAMBA	7.470	7.938	11006.3E6	4377.2E6	788.054	770.275
6) T	MCPD	7.647	8.035	654.3E6	134.0E6	81.330m	75.818m
7) T	MCPA	7.797	8.282	778.6E6	192.2E6	78.161	68.516
8) T	DICHLORPROP	8.178	8.657	2463.7E6	1001.3E6	776.593	750.008m
9) T	2,4-D	8.414	8.999	2361.6E6	1116.4E6	820.916	782.203m
10) T	Pentachlo...	8.710	9.508	40192.9E6	28331.7E6	837.136	813.465
11) T	2,4,5-TP ...	9.286	9.893	14816.0E6	10275.7E6	826.613	783.449
12) T	2,4,5-T	9.582	10.322	12701.6E6	8986.1E6	866.928	758.977
13) T	2,4-DB	10.159	10.891	1592.3E6	650.0E6	828.571	673.663
14) T	DINOSEB	11.356	11.266	10346.9E6	7420.4E6	816.099	784.651
15) T	Picloram	11.185	12.390	9672.5E6	11767.8E6	740.954m	671.306
16) T	DCPA	11.654	12.307	19396.1E6	16874.3E6	851.479m	856.545m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090325\
Data File : PS031620.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2025 22:45
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/04/2025
Supervised By :mohammad ahmed 09/05/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 02:28:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080725.M
Quant Title : 8080.M
QLast Update : Fri Aug 08 03:17:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

