

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090420\
 Data File : PS012133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2020 12:18
 Operator : DD\AJ
 Sample : L3876-13MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 VNJ-215MSD

Manual Integrations
 APPROVED

Ankita
 9/8/2020 11:00:39 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 13:57:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.913	174.3E6	77201524	159.834	144.359
Target Compounds							
1) T	Dalapon	3.035	2.628	202.9E6	107.1E6	162.912	141.686m
2) T	3,5-DICHL...	6.432	6.065	169.0E6	93928552	116.587	114.182m
3) T	4-Nitroph...	7.049	6.522	79595379	53252208	135.127m	146.053m
5) T	DICAMBA	7.408	7.081	520.9E6	278.0E6	127.180	128.674
6) T	MCPD	7.572	7.172	41673681	21999271	13.043	12.447
7) T	MCPA	7.721	7.384	66713706	78079394	14.667	28.495 #
8) T	DICHLORPROP	8.099	7.706	113.9E6	73025190	113.408	129.331m
9) T	2,4-D	8.329	8.001	141.4E6	110.9E6	125.209	150.955
10) T	Pentachlo...	8.639	8.429	2748.3E6	1529.6E6	191.100	225.120
11) T	2,4,5-TP ...	9.187	8.790	714.7E6	412.1E6	122.914m	140.945
12) T	2,4,5-T	9.473	9.169	637.6E6	365.3E6	107.604	123.406
13) T	2,4-DB	10.025	9.675	84675565	62769866	106.567	150.087m#
14) T	DINOSEB	11.183	10.024	119.2E6	155.2E6	25.394	72.986 #
15) T	Picloram	11.006	10.991	1178.6E6	747.3E6	150.804	164.877
16) T	DCPA	11.450	10.936	1019.1E6	562.4E6	134.030	152.946

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090420\
Data File : PS012133.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2020 12:18
Operator : DD\AJ
Sample : L3876-13MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
VNJ-215MSD

Manual Integrations
APPROVED

Ankita
9/8/2020 11:00:39 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 13:57:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

