

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090519\
 Data File : PS006563.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Sep 2019 09:14
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 9/6/2019 9:04:25 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 03:35:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.487	5.210	688.7E6	360.0E6	797.793	731.826
Target Compounds							
1) T	Dalapon	2.055	1.757	699.7E6	379.2E6	642.167	604.246
2) T	3,5-DICHL...	4.854	4.513	758.5E6	398.8E6	703.555	658.463
3) T	4-Nitroph...	5.340	4.895	259.8E6	158.4E6	692.162	639.283
5) T	DICAMBA	5.639	5.338	2408.9E6	1268.3E6	750.063	687.601
6) T	MCPD	5.791	5.422	156.1E6	86202114	69.353	69.681
7) T	MCPA	5.924	5.600	227.9E6	128.5E6	72.192	62.866
8) T	DICHLORPROP	6.267	5.877	554.9E6	284.9E6	731.730m	640.754
9) T	2,4-D	6.482	6.148	628.8E6	382.4E6	771.713	695.653
10) T	Pentachlo...	6.744	6.515	8962.8E6	4795.4E6	758.215	709.390
11) T	2,4,5-TP ...	7.278	6.878	3472.0E6	1825.1E6	761.104	701.406
12) T	2,4,5-T	7.550	7.237	3815.2E6	1802.2E6	771.590	724.941m
13) T	2,4-DB	8.080	7.724	487.5E6	234.0E6	782.833	703.175
14) T	DINOSEB	9.171	8.034	2754.4E6	1344.0E6	752.411	699.687
15) T	Picloram	9.002	8.971	5788.2E6	3123.7E6	777.919	692.479
16) T	DCPA	9.443	8.929	4542.8E6	2522.7E6	783.844	723.419m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090519\
Data File : PS006563.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2019 09:14
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
9/6/2019 9:04:25 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 03:35:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

