

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090523\
 Data File : PS024248.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Sep 2023 19:34
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/06/2023
 Supervised By : Ankita Jodhani 09/06/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 02:58:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.364	6.791	1503.8E6	524.2E6	754.822	780.472
Target Compounds							
1) T	Dalapon	2.483	2.549	1520.7E6	741.5E6	699.642	703.143
2) T	3,5-DICHL...	5.738	5.975	2115.4E6	765.8E6	695.909	734.433
3) T	4-Nitroph...	6.201	6.404	874.1E6	299.5E6	703.541	710.246
5) T	DICAMBA	6.506	6.954	6324.3E6	2167.1E6	725.497	752.415
6) T	MCPD	6.658	7.051	467.3E6	146.0E6	75.046	76.217
7) T	MCPA	6.781	7.258	631.1E6	221.1E6	72.710	75.638
8) T	DICHLORPROP	7.094	7.570	1615.2E6	534.7E6	710.479	735.956
9) T	2,4-D	7.291	7.854	1899.6E6	656.5E6	714.158	736.903
10) T	Pentachlo...	7.521	8.280	23032.2E6	8014.5E6	736.497	756.461
11) T	2,4,5-TP ...	8.042	8.638	9733.7E6	3132.6E6	728.202	754.993
12) T	2,4,5-T	8.297	9.008	9045.9E6	3014.9E6	730.095	751.278
13) T	2,4-DB	8.801	9.512	1539.4E6	376.3E6	714.126	818.677
14) T	DINOSEB	9.856	9.840	6757.7E6	2235.2E6	699.377m	747.735
15) T	Picloram	9.681	10.793	11459.8E6	4243.7E6	701.713	714.582
16) T	DCPA	10.121	10.763	11790.4E6	3964.5E6	728.072	772.808

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090523\
Data File : PS024248.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2023 19:34
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/06/2023
Supervised By : Ankita Jodhani 09/06/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 02:58:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

