

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090619\
 Data File : PS006581.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2019 17:23
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 9/9/2019 11:23:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 02:56:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
 Quant Title : 8080.M
 QLast Update : Wed Aug 14 15:09:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.483	5.213	628.2E6	379.3E6	727.732m	771.234m
Target Compounds							
1) T	Dalapon	2.051	1.760	716.9E6	417.1E6	657.890	664.715
2) T	3,5-DICHL...	4.850	4.516	838.2E6	445.0E6	777.423	734.729
3) T	4-Nitroph...	5.336	4.899	256.9E6	181.0E6	684.332	730.528
5) T	DICAMBA	5.636	5.341	2531.6E6	1345.8E6	788.253	729.625m
6) T	MCPD	5.787	5.425	179.9E6	90795296	79.955m	73.393
7) T	MCPA	5.921	5.603	255.0E6	148.4E6	80.767	72.621
8) T	DICHLORPROP	6.264	5.881	612.1E6	329.4E6	807.181	740.954
9) T	2,4-D	6.479	6.152	573.5E6	407.6E6	703.848m	741.534
10) T	Pentachlo...	6.741	6.518	9699.6E6	5175.1E6	820.544	765.560
11) T	2,4,5-TP ...	7.275	6.881	3716.5E6	1980.8E6	814.706m	761.229
12) T	2,4,5-T	7.547	7.240	4187.9E6	1899.2E6	846.974	763.955m
13) T	2,4-DB	8.078	7.727	532.0E6	258.4E6	854.159	776.293
14) T	DINOSEB	9.167	8.037	3030.5E6	1468.8E6	827.829	764.670
15) T	Picloram	9.001	8.975	5578.2E6	2973.0E6	749.696	659.079
16) T	DCPA	9.440	8.932	4967.2E6	2747.3E6	857.069	787.831m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090619\
Data File : PS006581.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2019 17:23
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
9/9/2019 11:23:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 02:56:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081419.M
Quant Title : 8080.M
QLast Update : Wed Aug 14 15:09:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

