

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090623\
 Data File : PS024257.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2023 19:20
 Operator : AR\AJ
 Sample : 04274-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 RT2683MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/07/2023
 Supervised By : Ankita Jodhani 09/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 20:52:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.364	6.791	1008.1E6	314.3E6	506.012	467.932
Target Compounds							
1) T	Dalapon	2.486	2.550	845.0E6	420.7E6	388.779	398.928
2) T	3,5-DICHL...	5.737	5.975	1554.5E6	498.4E6	511.391	477.961
5) T	DICAMBA	6.505	6.954	4158.1E6	1352.9E6	477.006	469.738
6) T	MCP	6.656	7.049	350.9E6	95449818	56.357	49.826
7) T	MCPA	6.778	7.255	391.8E6	132.0E6	45.135	45.146
8) T	DICHLORPROP	7.094	7.569	1090.2E6	361.2E6	479.552	497.082
9) T	2,4-D	7.290	7.853	1347.8E6	477.1E6	506.721	535.568
10) T	Pentachlo...	7.520	8.279	13585.1E6	4373.4E6	434.408	412.788
11) T	2,4,5-TP ...	8.041	8.637	6777.0E6	2653.0E6	507.001	639.404m#
12) T	2,4,5-T	8.296	9.008	6418.2E6	2036.0E6	518.012	507.349
13) T	2,4-DB	8.802	9.511	828.8E6	230.9E6	384.495	502.472 #
14) T	DINOSEB	9.856	9.840	2946.6E6	907.2E6	304.957	303.470
15) T	Picloram	9.680	10.793	7032.4E6	2558.5E6	430.612	430.819
16) T	DCPA	10.121	10.763	8366.0E6	2613.1E6	516.613	509.375

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090623\
Data File : PS024257.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2023 19:20
Operator : AR\AJ
Sample : 04274-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
RT2683MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/07/2023
Supervised By : Ankita Jodhani 09/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 20:52:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

