

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090723\
 Data File : PS024273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2023 18:56
 Operator : AR\AJ
 Sample : 04256-05MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MW-3MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/08/2023
 Supervised By : Ankita Jodhani 09/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 05:51:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.363	6.790	918.9E6	299.1E6	461.230	445.310
Target Compounds							
1) T	Dalapon	2.484	2.551	576.2E6	320.5E6	265.081	303.889
2) T	3,5-DICHL...	5.737	5.974	1182.1E6	378.0E6	388.864	362.495
3) T	4-Nitroph...	6.201	6.404	443.0E6	151.5E6	356.603	359.118
5) T	DICAMBA	6.505	6.952	3413.4E6	1045.4E6	391.574	362.979
6) T	MCPD	6.654	7.047	216.4E6	73317131	34.744m	38.272m
7) T	MCPA	6.778	7.254	329.1E6	97799848	37.917	33.458
8) T	DICHLORPROP	7.094	7.569	806.7E6	285.4E6	354.834	392.819
9) T	2,4-D	7.290	7.853	1047.1E6	357.0E6	393.670	400.736
10) T	Pentachlo...	7.520	8.278	7124.9E6	2233.9E6	227.831	210.849
11) T	2,4,5-TP ...	8.042	8.637	4239.0E6	1328.0E6	317.132	320.063
12) T	2,4,5-T	8.296	9.007	4476.0E6	1435.0E6	361.257	357.585
13) T	2,4-DB	8.803	9.511	640.9E6	168.3E6	297.334	366.193
15) T	Picloram	9.680	10.792	5834.4E6	2113.2E6	357.258	355.832
16) T	DCPA	10.122	10.763	7129.2E6	2137.9E6	440.237	416.748

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090723\
Data File : PS024273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2023 18:56
Operator : AR\AJ
Sample : 04256-05MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MW-3MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/08/2023
Supervised By :Ankita Jodhani 09/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 05:51:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

