

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090820\
 Data File : PS012152.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2020 17:56
 Operator : DD\AJ
 Sample : PB131479BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB131479BS

Manual Integrations
 APPROVED

Ankita
 9/9/2020 10:43:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 09 02:36:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.913	522.2E6	292.0E6	478.715	546.012m
Target Compounds							
1) T	Dalapon	3.034	2.627	572.8E6	308.1E6	459.844	407.632
2) T	3,5-DICHL...	6.431	6.065	727.3E6	398.6E6	501.710	484.553
3) T	4-Nitroph...	7.048	6.522	194.6E6	112.4E6	330.283	308.364
5) T	DICAMBA	7.406	7.080	1959.4E6	1068.4E6	478.364	494.503
6) T	MCPD	7.574	7.175	155.8E6	86709828	48.749	49.060
7) T	MCPA	7.722	7.390	210.9E6	145.6E6	46.371	53.136
8) T	DICHLORPROP	8.097	7.706	504.4E6	297.3E6	502.087	526.457
9) T	2,4-D	8.328	8.001	624.7E6	404.6E6	553.292	550.679
10) T	Pentachlo...	8.638	8.429	7796.6E6	3785.7E6	542.126	557.148
11) T	2,4,5-TP ...	9.186	8.790	3094.8E6	1609.4E6	532.235	550.472
12) T	2,4,5-T	9.472	9.170	2969.4E6	1548.5E6	501.100	523.113
13) T	2,4-DB	10.025	9.677	323.6E6	211.5E6	407.323	505.635
14) T	DINOSEB	11.183	10.008	2369.3E6	1128.1E6	504.783	530.656
15) T	Picloram	11.009	10.996	2494.9E6	1664.1E6	319.222	367.150
16) T	DCPA	11.450	10.936	2405.6E6	1482.4E6	316.380	403.174 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090820\
Data File : PS012152.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2020 17:56
Operator : DD\AJ
Sample : PB131479BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB131479BS

Manual Integrations
APPROVED

Ankita
9/9/2020 10:43:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 09 02:36:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

