

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090919\
 Data File : PS006614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2019 11:30
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 9/10/2019 12:24:18 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 06:16:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 10 06:14:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.483	5.207	596.2E6	282.3E6	767.445	661.012
Target Compounds							
1) T	Dalapon	2.048	1.754	753.2E6	346.4E6	770.623	677.757
2) T	3,5-DICHL...	4.849	4.509	608.1E6	359.6E6	614.205	664.470
3) T	4-Nitroph...	5.335	4.892	247.8E6	147.3E6	698.292	665.664
5) T	DICAMBA	5.637	5.336	2110.6E6	1150.3E6	732.118	717.518
6) T	MCPD	5.789	5.419	145.8E6	73222881	76.659	69.188
7) T	MCPA	5.922	5.597	273.8E6	112.7E6	96.867	64.271 #
8) T	DICHLORPROP	6.265	5.875	547.4E6	269.5E6	769.745	692.470
9) T	2,4-D	6.479	6.145	685.3E6	343.7E6	887.028	720.501
10) T	Pentachlo...	6.742	6.513	8271.0E6	4382.0E6	742.470	727.308
11) T	2,4,5-TP ...	7.276	6.875	3220.5E6	1556.3E6	748.434	683.676
12) T	2,4,5-T	7.549	7.270	3495.9E6	1619.9E6	750.245	751.310
13) T	2,4-DB	8.079	7.721	452.4E6	212.7E6	743.347	773.592
14) T	DINOSEB	9.169	8.032	2550.3E6	1225.3E6	732.675	731.369
15) T	Picloram	9.000	8.969	5170.8E6	2778.2E6	766.537	750.737
16) T	DCPA	9.441	8.969	4237.4E6	2778.2E6	750.096	881.901

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090919\
Data File : PS006614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2019 11:30
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
9/10/2019 12:24:18 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 06:16:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
Quant Title : 8080.M
QLast Update : Tue Sep 10 06:14:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

