

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090920\
 Data File : PS012173.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2020 08:54
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 09 10:46:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.913	819.8E6	426.0E6	751.591	796.557
Target Compounds							
1) T	Dalapon	3.035	2.625	810.9E6	444.1E6	650.962	587.600
2) T	3,5-DICHL...	6.431	6.065	1069.9E6	602.7E6	738.007	732.602
3) T	4-Nitroph...	7.048	6.522	365.6E6	233.1E6	620.743	639.356
5) T	DICAMBA	7.407	7.081	3157.6E6	1635.6E6	770.899	757.036
6) T	MCPD	7.576	7.178	240.0E6	132.1E6	75.111	74.730
7) T	MCPA	7.726	7.393	341.9E6	209.6E6	75.173	76.485
8) T	DICHLORPROP	8.098	7.706	751.8E6	430.5E6	748.318	762.491
9) T	2,4-D	8.328	8.001	818.3E6	552.2E6	724.751	751.606
10) T	Pentachlo...	8.638	8.429	10971.0E6	5057.7E6	762.850	744.350
11) T	2,4,5-TP ...	9.186	8.790	4383.7E6	2231.3E6	753.892	763.174
12) T	2,4,5-T	9.472	9.169	4487.5E6	2214.6E6	757.275	748.170
13) T	2,4-DB	10.024	9.676	535.0E6	307.2E6	673.321	734.604
14) T	DINOSEB	11.183	10.008	3481.9E6	1638.1E6	741.836	770.515
15) T	Picloram	11.008	10.994	4500.1E6	2744.0E6	575.781	605.418
16) T	DCPA	11.448	10.935	5458.0E6	2822.7E6	717.838	767.692

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090920\
Data File : PS012173.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2020 08:54
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 09 10:46:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

