

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090922\
 Data File : PS020495.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2022 12:17
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/12/2022
 Supervised By :Ankita Jodhani 09/12/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 09 23:19:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090222.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 03 01:06:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.178	6.753	2542.7E6	866.0E6	721.389	814.118m
Target Compounds							
1) T	Dalapon	2.996	2.506	3429.3E6	999.6E6	701.195	699.631
2) T	3,5-DICHL...	6.394	5.940	3221.9E6	1235.1E6	675.662	733.332
3) T	4-Nitroph...	7.009	6.376	1278.1E6	518.5E6	827.079	823.975
5) T	DICAMBA	7.367	6.916	9366.3E6	3578.9E6	665.797	747.818
6) T	MCPD	7.538	7.015	711.8E6	274.9E6	66.748	74.775
7) T	MCPA	7.687	7.227	1004.5E6	385.6E6	65.209	71.091
8) T	DICHLORPROP	8.058	7.535	2524.1E6	928.2E6	763.692	742.712
9) T	2,4-D	8.289	7.828	2977.1E6	1177.3E6	634.588	775.285
10) T	Pentachlo...	8.593	8.250	31214.9E6	12640.4E6	656.800	743.134
11) T	2,4,5-TP ...	9.144	8.611	13384.9E6	5444.2E6	630.815	752.001
12) T	2,4,5-T	9.429	8.988	14120.1E6	5305.7E6	608.886	783.642 #
13) T	2,4-DB	9.980	9.493	3239.8E6	708.4E6	744.049	782.373
14) T	DINOSEB	11.131	9.816	15135.2E6	3953.8E6	657.099	752.825
15) T	Picloram	10.958	10.796	30722.0E6	7910.4E6	741.918	770.253
16) T	DCPA	11.400	10.743	23030.2E6	6806.7E6	716.800	764.213

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS090922\
Data File : PS020495.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2022 12:17
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/12/2022
Supervised By : Ankita Jodhani 09/12/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 09 23:19:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090222.M
Quant Title : 8080.M
QLast Update : Sat Sep 03 01:06:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

