

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091020\
 Data File : PS012208.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2020 18:26
 Operator : DD\AJ
 Sample : L3870-05MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SU-04-090820MSD

Manual Integrations
 APPROVED

Ankita
 9/11/2020 11:35:08 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 18:58:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
 Quant Title : 8080.M
 QLast Update : Tue Aug 11 15:43:16 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.215	6.912	259.3E6	154.2E6	237.723	288.330
Target Compounds							
1) T	Dalapon	3.038	2.632	535.5E6	291.1E6	429.890	385.147
2) T	3,5-DICHL...	6.430	6.064	635.7E6	363.2E6	438.502	441.563m
5) T	DICAMBA	7.406	7.081	1918.1E6	1051.7E6	468.291	486.776
6) T	MCP	7.573	7.175	158.7E6	80578091	49.660	45.590
7) T	MCPA	7.721	7.390	221.2E6	132.7E6	48.624m	48.434m
8) T	DICHLORPROP	8.097	7.705	512.5E6	378.5E6	510.113	670.373 #
9) T	2,4-D	8.326	8.000	648.9E6	415.6E6	574.734	565.697
10) T	Pentachlo...	8.637	8.428	4155.6E6	2130.6E6	288.957	313.563
11) T	2,4,5-TP ...	9.185	8.784	3035.0E6	3050.4E6	521.942	1043.348 #
12) T	2,4,5-T	9.470	9.168	2939.2E6	1543.3E6	496.000	521.370
13) T	2,4-DB	10.023	9.675	314.5E6	210.8E6	395.855	503.960 #
14) T	DINOSEB	11.182	10.007	952.1E6	457.1E6	202.844	215.017
15) T	Picloram	11.005	10.992	3800.2E6	2356.4E6	486.234	519.892
16) T	DCPA	11.447	10.935	2342.9E6	1289.8E6	308.135	350.795

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091020\
Data File : PS012208.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2020 18:26
Operator : DD\AJ
Sample : L3870-05MSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SU-04-090820MSD

Manual Integrations
APPROVED

Ankita
9/11/2020 11:35:08 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 18:58:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS081120.M
Quant Title : 8080.M
QLast Update : Tue Aug 11 15:43:16 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

