

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091125\
 Data File : PS031705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2025 18:00
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 01:47:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 09 13:56:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.287	7.744	2886.9E6	687.4E6	801.564	756.274
Target Compounds							
1) T	Dalapon	2.674	2.693	4075.0E6	1987.5E6	704.933	705.376
2) T	3,5-DICHL...	6.455	6.697	3943.5E6	1109.8E6	739.398	727.725
3) T	4-Nitroph...	7.099	7.297	896.5E6	840.7E6	812.614	665.732
5) T	DICAMBA	7.471	7.940	11794.1E6	4769.9E6	756.189	733.062
6) T	MCPD	7.648	8.037	681.9E6	144.4E6	76.284	76.678
7) T	MCPA	7.797	8.285	850.2E6	215.4E6	76.520	76.895
8) T	DICHLORPROP	8.178	8.660	2635.0E6	1080.5E6	739.797	742.613
9) T	2,4-D	8.416	9.003	2587.4E6	1167.9E6	755.301	709.908
10) T	Pentachlo...	8.711	9.510	42576.7E6	30630.2E6	801.914	771.726
11) T	2,4,5-TP ...	9.288	9.895	15982.1E6	11177.7E6	762.005	742.489
12) T	2,4,5-T	9.585	10.324	14486.1E6	9887.5E6	791.672	734.747
13) T	2,4-DB	10.160	10.892	1844.9E6	755.1E6	790.875	705.697
14) T	DINOSEB	11.357	11.267	11876.5E6	8078.4E6	823.461	831.051
15) T	Picloram	11.188	12.389	14288.1E6	18167.4E6	793.296	767.098
16) T	DCPA	11.656	12.309	20862.5E6	17697.8E6	774.123	734.223

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091125\
Data File : PS031705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2025 18:00
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 01:47:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
Quant Title : 8080.M
QLast Update : Tue Sep 09 13:56:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

