

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091219\
 Data File : PS006679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2019 19:20
 Operator : SM\AJ
 Sample : PB123030BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB123030BS

Manual Integrations
 APPROVED

Ankita
 9/13/2019 3:53:46 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 04:51:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 10 06:14:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.479	5.209	589.5E6	319.4E6	758.753	748.081
Target Compounds							
1) T	Dalapon	2.045	1.754	734.4E6	412.3E6	751.342	806.739m
2) T	3,5-DICHL...	4.845	4.511	718.6E6	383.9E6	725.876m	709.266m
3) T	4-Nitroph...	5.339	4.897	114.0E6	54382719	321.354	245.687m
5) T	DICAMBA	5.632	5.337	2103.5E6	1158.1E6	729.643	722.339
6) T	MCPD	5.784	5.421	160.9E6	85370344	84.604m	80.667
7) T	MCPA	5.917	5.598	212.3E6	149.9E6	75.130m	85.483
8) T	DICHLORPROP	6.259	5.875	555.3E6	301.7E6	780.893	775.271
9) T	2,4-D	6.473	6.145	694.3E6	427.2E6	898.690	895.690
10) T	Pentachlo...	6.736	6.513	8747.1E6	4725.8E6	785.207	784.361
11) T	2,4,5-TP ...	7.270	6.875	3396.2E6	1857.8E6	789.261	816.094m
12) T	2,4,5-T	7.542	7.235	3405.7E6	1761.2E6	730.884	816.841m
13) T	2,4-DB	8.072	7.721	486.2E6	226.1E6	799.024	822.194
14) T	DINOSEB	9.163	8.032	2445.4E6	1254.2E6	702.543	748.620
15) T	Picloram	8.994	8.969	4359.7E6	2306.4E6	646.288	623.250m
16) T	DCPA	9.435	8.927	2960.5E6	1477.8E6	524.056	469.102m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091219\
Data File : PS006679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2019 19:20
Operator : SM\AJ
Sample : PB123030BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB123030BS

Manual Integrations
APPROVED

Ankita
9/13/2019 3:53:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 04:51:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
Quant Title : 8080.M
QLast Update : Tue Sep 10 06:14:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

