

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091318\
 Data File : PS001736.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2018 18:29
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 04:16:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 23 02:55:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.464	10.963	142.9E6	164.5E6	773.973m	812.031
Target Compounds							
1) T	Dalapon	2.933	2.395	183.1E6	176.2E6	728.376m	736.208m
2) T	3,5-DICHL...	9.886	9.112	196.5E6	190.6E6	733.600m	745.754
3) T	4-Nitroph...	11.055	10.108	87283859	74992136	707.942	711.477m
5) T	DICAMBA	11.776	11.249	520.3E6	543.5E6	744.822	774.902m
6) T	MCPD	12.116	11.475	38270779	37643903	76.061m	81.749m
7) T	MCPA	12.367	11.852	61726447	55685415	73.933m	76.404m
8) T	DICHLORPROP	12.966	12.373	139.0E6	142.3E6	735.853	763.843m
9) T	2,4-D	13.324	12.844	172.8E6	189.5E6	728.904	765.166m
10) T	Pentachlo...	13.706	13.395	1904.5E6	2156.0E6	763.748	797.268m
11) T	2,4,5-TP ...	14.577	14.012	715.7E6	847.4E6	766.961	797.303
12) T	2,4,5-T	14.973	14.580	725.5E6	694.0E6	766.609	803.320
13) T	2,4-DB	15.722	15.228	101.7E6	105.1E6	756.990	792.368m
14) T	DINOSEB	17.211	15.658	501.0E6	579.1E6	767.738m	783.045m
15) T	Picloram	16.972	16.839	1071.7E6	1068.4E6	765.157m	819.364m
16) T	DCPA	17.552	16.894	908.1E6	1184.7E6	800.330m	829.784m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091318\
Data File : PS001736.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2018 18:29
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 04:16:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
Quant Title : 8080.M
QLast Update : Thu Aug 23 02:55:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

