

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091319\
 Data File : PS006708.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2019 15:00
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 15:24:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 10 06:14:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.501	5.230	579.3E6	313.4E6	745.654	733.929
Target Compounds							
1) T	Dalapon	2.071	1.781	589.1E6	315.0E6	602.677	616.301
2) T	3,5-DICHL...	4.868	4.533	704.0E6	364.3E6	711.103	673.071
3) T	4-Nitroph...	5.355	4.916	241.5E6	147.3E6	680.462	665.482
5) T	DICAMBA	5.653	5.358	2032.4E6	1106.0E6	704.967	689.868
6) T	MCPD	5.804	5.441	134.8E6	76952302	70.884	72.712
7) T	MCPA	5.937	5.619	208.0E6	120.8E6	73.605	68.850
8) T	DICHLORPROP	6.281	5.897	501.2E6	268.3E6	704.806	689.376
9) T	2,4-D	6.497	6.168	560.8E6	328.4E6	725.776	688.437
10) T	Pentachlo...	6.757	6.534	8028.6E6	4214.0E6	720.702	699.425
11) T	2,4,5-TP ...	7.291	6.898	3059.6E6	1605.4E6	711.032	705.212
12) T	2,4,5-T	7.564	7.257	3362.1E6	1520.6E6	721.528	705.218
13) T	2,4-DB	8.094	7.743	440.2E6	199.1E6	723.368	724.135
14) T	DINOSEB	9.182	8.053	2556.5E6	1165.4E6	734.466	695.628
15) T	Picloram	9.015	8.991	5007.8E6	2520.5E6	742.365	681.094
16) T	DCPA	9.455	8.948	4178.4E6	2202.8E6	739.648	699.247

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

