

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091322\
 Data File : PS020509.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2022 22:01
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 02:55:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 14 02:51:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.177	6.759	2503.5E6	628.9E6	727.433	738.397
Target Compounds							
1) T	Dalapon	3.002	2.516	3340.9E6	811.6E6	665.646	669.356
2) T	3,5-DICHL...	6.395	5.945	3361.8E6	941.2E6	674.178	689.709
3) T	4-Nitroph...	7.019	6.388	1017.2E6	330.7E6	656.032	701.857
5) T	DICAMBA	7.363	6.919	9386.8E6	2715.8E6	686.465	705.472
6) T	MCPD	7.531	7.014	730.7E6	205.2E6	70.168	71.057
7) T	MCPA	7.679	7.225	1017.7E6	296.0E6	68.616	68.549
8) T	DICHLORPROP	8.056	7.539	2429.9E6	700.3E6	673.923	694.082
9) T	2,4-D	8.291	7.834	3179.4E6	894.6E6	684.029	704.362
10) T	Pentachlo...	8.588	8.252	33127.7E6	9970.1E6	718.239	703.547
11) T	2,4,5-TP ...	9.140	8.614	13969.6E6	4198.9E6	683.442	709.400
12) T	2,4,5-T	9.428	8.993	15089.1E6	3915.5E6	681.367	718.505
13) T	2,4-DB	9.982	9.497	2578.2E6	506.7E6	709.983	708.279
14) T	DINOSEB	11.124	9.817	11903.8E6	3000.8E6	648.166	699.848
15) T	Picloram	10.962	10.803	21368.3E6	5840.1E6	710.070	725.212
16) T	DCPA	11.394	10.744	17525.4E6	5253.7E6	684.102	712.318

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091322\
Data File : PS020509.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2022 22:01
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 02:55:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
Quant Title : 8080.M
QLast Update : Wed Sep 14 02:51:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

