

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091322\
 Data File : PS020526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 06:30
 Operator : AJ\MA
 Sample : N4539-03MS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MBD-WC-S-SB-1-2-0-5MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/14/2022
 Supervised By : Ankita Jodhani 09/14/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 07:17:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 14 02:51:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.174	6.756	576.6E6	179.7E6	167.534	210.970 #
Target Compounds							
1) T	Dalapon	2.998	2.519	1805.8E6	531.4E6	359.793m	438.297
2) T	3,5-DICHL...	6.390	5.943	1594.5E6	409.4E6	319.751	299.997
3) T	4-Nitroph...	7.058	6.427	165.5E6	85250835	213.173m	180.917
5) T	DICAMBA	7.361	6.917	5530.9E6	1587.4E6	404.482	412.363
6) T	MCPD	7.527	7.011	469.4E6	120.7E6	45.078	41.779
7) T	MCPA	7.674	7.222	1168.0E6	230.4E6	78.752m	53.368 #
8) T	DICHLORPROP	8.051	7.536	1548.6E6	472.6E6	429.502	468.352
9) T	2,4-D	8.280	7.828	1953.2E6	623.4E6	420.220m	490.844
10) T	Pentachlo...	8.586	8.251	9712.3E6	2779.3E6	210.572	196.122
11) T	2,4,5-TP ...	9.135	8.606	9383.8E6	6151.2E6	459.088	1039.244m#
12) T	2,4,5-T	9.422	8.991	9097.5E6	2578.9E6	410.812	473.232
13) T	2,4-DB	9.973	9.494	1243.1E6	376.9E6	342.335	526.818 #
14) T	DINOSEB	11.122	9.817	2934.7E6	655.1E6	159.795	152.790
15) T	Picloram	10.952	10.798	13728.7E6	3163.5E6	456.204	392.837
16) T	DCPA	11.392	10.743	10237.8E6	3074.1E6	399.632	416.792

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091322\
Data File : PS020526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 06:30
Operator : AJ\MA
Sample : N4539-03MS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MBD-WC-S-SB-1-2-0-5MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/14/2022
Supervised By :Ankita Jodhani 09/14/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 07:17:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
Quant Title : 8080.M
QLast Update : Wed Sep 14 02:51:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

