

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091418\
 Data File : PS001751.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2018 22:14
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 9/17/2018 12:48:25 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 11:28:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 23 02:55:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.479	10.961	143.3E6	159.6E6	776.416	788.006
Target Compounds							
1) T	Dalapon	2.949	2.395	180.4E6	173.7E6	717.633	725.688m
2) T	3,5-DICHL...	9.902	9.112	197.8E6	184.7E6	738.688	722.677
3) T	4-Nitroph...	11.069	10.107	75575639	68119754	612.979	646.276m
5) T	DICAMBA	11.792	11.246	547.9E6	550.6E6	784.458	785.029m
6) T	MCPD	12.131	11.474	35120279	34146747	69.799	74.155m
7) T	MCPA	12.382	11.850	58855774	50672867	70.494m	69.527m
8) T	DICHLORPROP	12.981	12.371	141.5E6	138.8E6	749.144	745.260m
9) T	2,4-D	13.338	12.842	175.6E6	181.7E6	740.505	733.429m
10) T	Pentachlo...	13.721	13.394	1977.2E6	2163.6E6	792.891	800.097
11) T	2,4,5-TP ...	14.591	14.010	748.3E6	854.2E6	801.846m	803.636
12) T	2,4,5-T	14.988	14.578	745.5E6	692.0E6	787.739m	801.009
13) T	2,4-DB	15.736	15.226	108.0E6	103.7E6	803.786	782.081m
14) T	DINOSEB	17.225	15.657	536.0E6	580.1E6	821.453	784.357m
15) T	Picloram	16.986	16.892	1125.5E6	1098.4E6	803.597	842.327m
16) T	DCPA	17.566	16.892	965.0E6	1117.7E6	850.497m	782.825m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091418\
Data File : PS001751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2018 22:14
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
9/17/2018 12:48:25 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 11:28:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
Quant Title : 8080.M
QLast Update : Thu Aug 23 02:55:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

