

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091420\
 Data File : PS012240.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2020 17:36
 Operator : DD\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 05:53:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 15 05:52:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.913	555.2E6	270.4E6	520.056	518.626
Target Compounds							
1) T	Dalapon	3.032	2.620	573.9E6	316.5E6	468.514	472.104
2) T	3,5-DICHL...	6.432	6.065	704.5E6	391.7E6	486.679	489.704
3) T	4-Nitroph...	7.050	6.523	245.8E6	150.7E6	465.567	470.070
5) T	DICAMBA	7.407	7.080	2058.6E6	1071.9E6	491.898	500.117
6) T	MCPD	7.574	7.175	153.4E6	82404662	47.195	48.017
7) T	MCPA	7.723	7.389	219.9E6	133.4E6	47.071	48.603
8) T	DICHLORPROP	8.099	7.706	491.9E6	274.6E6	488.623	495.099
9) T	2,4-D	8.330	8.003	547.3E6	353.2E6	486.335	493.045
10) T	Pentachlo...	8.638	8.429	7518.9E6	3530.7E6	516.226	531.212
11) T	2,4,5-TP ...	9.188	8.790	2933.3E6	1463.2E6	502.304	510.400
12) T	2,4,5-T	9.474	9.171	3039.0E6	1454.9E6	502.522	506.041
13) T	2,4-DB	10.028	9.678	368.0E6	197.2E6	481.862	491.138
14) T	DINOSEB	11.183	10.008	2303.3E6	1021.1E6	495.333	494.939
15) T	Picloram	11.013	10.999	3560.1E6	2131.8E6	480.293	492.823
16) T	DCPA	11.449	10.935	3756.0E6	1870.3E6	512.288	520.352

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091420\
Data File : PS012240.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2020 17:36
Operator : DD\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 05:53:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
Quant Title : 8080.M
QLast Update : Tue Sep 15 05:52:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

