

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091422\
 Data File : PS020560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2022 07:53
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 08:16:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 14 02:51:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.173	6.756	2740.2E6	698.2E6	796.215	819.734
Target Compounds							
1) T	Dalapon	3.002	2.518	3501.8E6	855.4E6	697.696	705.451
2) T	3,5-DICHL...	6.390	5.942	3703.3E6	1025.3E6	742.654	751.323
3) T	4-Nitroph...	7.007	6.380	1264.5E6	372.8E6	784.604	791.173
5) T	DICAMBA	7.361	6.917	10382.8E6	2957.0E6	759.304	768.134
6) T	MCPD	7.530	7.014	825.6E6	224.0E6	79.284	77.550
7) T	MCPA	7.679	7.226	1140.1E6	319.2E6	76.870	73.926
8) T	DICHLORPROP	8.051	7.537	2660.4E6	763.6E6	737.855	756.785
9) T	2,4-D	8.282	7.830	3362.8E6	964.0E6	723.479	759.022
10) T	Pentachlo...	8.586	8.251	35783.1E6	10739.5E6	775.810	757.838
11) T	2,4,5-TP ...	9.136	8.612	15070.9E6	4532.6E6	737.321	765.780
12) T	2,4,5-T	9.421	8.989	15825.4E6	4361.9E6	714.615	800.433
13) T	2,4-DB	9.973	9.493	2939.0E6	570.7E6	809.331	797.708
14) T	DINOSEB	11.121	9.816	11980.5E6	3217.9E6	652.344	750.474
15) T	Picloram	10.950	10.797	23235.8E6	6061.9E6	772.128	752.751
16) T	DCPA	11.391	10.743	18746.8E6	5685.2E6	731.776	770.820

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091422\
Data File : PS020560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2022 07:53
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 08:16:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
Quant Title : 8080.M
QLast Update : Wed Sep 14 02:51:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

