

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091423\
 Data File : PS024303.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2023 18:53
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/15/2023
 Supervised By : Ankita Jodhani 09/15/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 02:00:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.363	6.789	1523.6E6	517.5E6	764.755	770.521
Target Compounds							
1) T	Dalapon	2.483	2.547	1468.7E6	690.5E6	675.714	654.797
2) T	3,5-DICHL...	5.737	5.972	2145.4E6	756.6E6	705.777	725.665
3) T	4-Nitroph...	6.200	6.401	881.4E6	289.9E6	709.415	687.456
5) T	DICAMBA	6.505	6.951	6419.1E6	2139.2E6	736.378	742.732
6) T	MCPD	6.657	7.047	466.3E6	150.6E6	74.880	78.624m
7) T	MCPA	6.780	7.255	639.7E6	211.6E6	73.701	72.390
8) T	DICHLORPROP	7.093	7.566	1639.4E6	518.7E6	721.125	713.964
9) T	2,4-D	7.290	7.850	1933.7E6	645.5E6	726.971	724.524
10) T	Pentachlo...	7.520	8.276	23546.3E6	7886.3E6	752.936	744.363
11) T	2,4,5-TP ...	8.041	8.634	9897.9E6	3071.4E6	740.486	740.234
12) T	2,4,5-T	8.297	9.004	9189.3E6	2964.5E6	741.665	738.721
13) T	2,4-DB	8.801	9.508	1546.9E6	369.7E6	717.595	804.395
14) T	DINOSEB	9.856	9.836	7063.9E6	2199.7E6	731.072	735.859
15) T	Picloram	9.680	10.789	11806.7E6	4332.5E6	722.956	729.528
16) T	DCPA	10.121	10.758	11998.2E6	3834.9E6	740.908	747.554

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091423\
Data File : PS024303.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2023 18:53
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/15/2023
Supervised By : Ankita Jodhani 09/15/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 02:00:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

