

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091424\
 Data File : PS027679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2024 14:18
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 16 01:29:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 14 02:36:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.109	7.633	1887.8E6	710.8E6	758.877	717.188
Target Compounds							
1) T	Dalapon	2.554	2.621	2720.5E6	1069.1E6	656.642	655.113
2) T	3,5-DICHL...	6.299	6.604	2579.0E6	1130.7E6	706.020	701.665
3) T	4-Nitroph...	6.905	7.167	1168.5E6	498.2E6	690.155	695.082
5) T	DICAMBA	7.288	7.827	7315.9E6	3052.4E6	723.973	692.717
6) T	MCP	7.468	7.931	508.9E6	264.1E6	73.820	80.444
7) T	MCPA	7.614	8.171	730.6E6	395.4E6	72.122	74.690
8) T	DICHLORPROP	7.975	8.534	1887.2E6	767.1E6	711.908	686.570
9) T	2,4-D	8.198	8.860	2163.5E6	756.8E6	705.816	678.769
10) T	Pentachlo...	8.485	9.376	26531.1E6	12417.5E6	729.455	751.104
11) T	2,4,5-TP ...	9.052	9.753	10727.0E6	4234.4E6	732.965	706.222
12) T	2,4,5-T	9.338	10.169	10836.2E6	3643.9E6	723.966	697.147
13) T	2,4-DB	9.902	10.731	1662.0E6	427.2E6	715.675	694.284
14) T	DINOSEB	11.080	11.107	7120.5E6	2966.1E6	728.004	705.282
15) T	Picloram	10.897	12.188	13029.7E6	3545.9E6	696.490	763.006
16) T	DCPA	11.379	12.141	12290.6E6	4513.1E6	740.515	700.231

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091424\
Data File : PS027679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2024 14:18
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 16 01:29:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
Quant Title : 8080.M
QLast Update : Sat Sep 14 02:36:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

