

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091520\
 Data File : PS012272.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2020 01:13
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 16 06:33:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 15 06:03:04 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.912	836.8E6	398.2E6	772.052	754.801
Target Compounds							
1) T	Dalapon	3.032	2.621	829.4E6	445.6E6	675.623	669.394
2) T	3,5-DICHL...	6.431	6.064	1050.2E6	566.0E6	716.287	699.585
3) T	4-Nitroph...	7.048	6.520	380.7E6	223.1E6	714.840	697.421
5) T	DICAMBA	7.407	7.079	3083.3E6	1538.9E6	732.795	713.514
6) T	MCPD	7.576	7.176	241.1E6	123.9E6	74.819	72.015
7) T	MCPA	7.725	7.391	343.6E6	192.8E6	73.547	68.598
8) T	DICHLORPROP	8.098	7.705	744.5E6	398.9E6	727.977	704.957
9) T	2,4-D	8.328	8.000	829.7E6	513.4E6	728.507	705.828
10) T	Pentachlo...	8.638	8.428	10904.2E6	4833.6E6	741.310	713.682
11) T	2,4,5-TP ...	9.186	8.788	4368.7E6	2090.3E6	742.278	722.256
12) T	2,4,5-T	9.472	9.168	4576.5E6	2101.0E6	746.205	729.475
13) T	2,4-DB	10.026	9.676	575.5E6	293.7E6	752.157	721.883
14) T	DINOSEB	11.183	10.006	3478.4E6	1527.6E6	740.701	735.523
15) T	Picloram	11.009	10.994	5329.1E6	2979.8E6	723.763	702.871
16) T	DCPA	11.448	10.934	5535.0E6	2666.2E6	747.927	733.530

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091520\
Data File : PS012272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2020 01:13
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 16 06:33:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
Quant Title : 8080.M
QLast Update : Tue Sep 15 06:03:04 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

