

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091619\
 Data File : PS006745.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2019 16:16
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 9/18/2019 9:38:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 16 17:29:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 10 06:14:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.496	5.225	578.1E6	321.7E6	744.143	753.351
Target Compounds							
1) T	Dalapon	2.064	1.772	710.5E6	368.1E6	726.967	720.207
2) T	3,5-DICHL...	4.863	4.527	729.3E6	385.8E6	736.634	712.783
3) T	4-Nitroph...	5.350	4.911	236.7E6	146.9E6	667.110	663.702
5) T	DICAMBA	5.648	5.352	2047.3E6	1129.0E6	710.154	704.185
6) T	MCPD	5.799	5.436	129.0E6	77479032	67.816	73.210
7) T	MCPA	5.932	5.613	204.2E6	118.2E6	72.247	67.409
8) T	DICHLORPROP	6.276	5.891	506.7E6	275.5E6	712.505	707.761
9) T	2,4-D	6.491	6.163	584.9E6	342.3E6	756.992	717.532
10) T	Pentachlo...	6.752	6.529	8916.6E6	4687.7E6	800.418	778.043
11) T	2,4,5-TP ...	7.287	6.892	3132.1E6	1671.9E6	727.892	734.444
12) T	2,4,5-T	7.559	7.251	3509.6E6	1572.8E6	753.193	729.422m
13) T	2,4-DB	8.089	7.737	431.5E6	198.7E6	709.088	722.571
14) T	DINOSB	9.178	8.047	2719.0E6	1293.1E6	781.150	771.825
15) T	Picloram	9.012	8.986	4534.4E6	2441.1E6	672.188	659.646
16) T	DCPA	9.450	8.942	4209.3E6	2328.6E6	745.123	739.187

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091619\
Data File : PS006745.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2019 16:16
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
9/18/2019 9:38:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 16 17:29:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
Quant Title : 8080.M
QLast Update : Tue Sep 10 06:14:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

