

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091621\
 Data File : PS016834.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2021 10:53
 Operator : DD\AJ
 Sample : PB139116BSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB139116BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 02:42:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS083121.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 01 10:13:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.180	6.761	697.4E6	267.1E6	464.381	515.226
Target Compounds							
1) T	Dalapon	3.012	2.516	953.1E6	292.6E6	423.020	435.207
2) T	3,5-DICHL...	6.399	5.954	863.2E6	379.8E6	405.460	478.146
3) T	4-Nitroph...	7.011	6.385	343.2E6	142.4E6	443.836	480.976
5) T	DICAMBA	7.369	6.923	2545.8E6	1099.8E6	447.314	498.218
6) T	MCPD	7.535	7.017	182.6E6	86412720	43.299	50.694
7) T	MCPA	7.683	7.229	266.3E6	130.0E6	45.463	48.516
8) T	DICHLORPROP	8.061	7.543	620.5E6	273.1E6	452.413	501.776
9) T	2,4-D	8.292	7.835	702.3E6	317.1E6	463.640	504.893
10) T	Pentachlo...	8.597	8.260	9388.7E6	3887.4E6	480.738	521.057
11) T	2,4,5-TP ...	9.147	8.621	3453.6E6	1520.9E6	480.998	517.371
12) T	2,4,5-T	9.433	8.998	3599.3E6	1483.4E6	503.006	527.876
13) T	2,4-DB	9.984	9.504	721.4E6	186.8E6	521.925	511.825
14) T	DINOSEB	11.136	9.829	2668.3E6	1075.5E6	489.358	512.907
15) T	Picloram	10.962	10.811	4854.8E6	1936.3E6	544.818	528.448
16) T	DCPA	11.404	10.759	5001.0E6	2148.1E6	502.048	495.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091621\
Data File : PS016834.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2021 10:53
Operator : DD\AJ
Sample : PB139116BSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB139116BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 02:42:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS083121.M
Quant Title : 8080.M
QLast Update : Wed Sep 01 10:13:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

