

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091624\
 Data File : PS027694.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2024 13:28
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/17/2024
 Supervised By :Ankita Jodhani 09/17/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 03:25:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 14 02:36:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.108	7.630	1829.8E6	686.6E6	735.563	692.764
Target Compounds							
1) T	Dalapon	2.553	2.625	2756.3E6	1049.6E6	665.288	643.183
2) T	3,5-DICHL...	6.299	6.602	2532.7E6	1057.0E6	693.339	655.947
3) T	4-Nitroph...	6.904	7.165	1151.7E6	478.9E6	680.209	668.152
5) T	DICAMBA	7.288	7.826	7095.7E6	2866.2E6	702.190	650.473
6) T	MCPD	7.467	7.928	484.7E6	242.4E6	70.309	73.850
7) T	MCPA	7.613	8.169	702.4E6	359.1E6	69.336	67.829
8) T	DICHLORPROP	7.974	8.532	1829.3E6	730.5E6	690.060	653.880
9) T	2,4-D	8.197	8.857	2116.2E6	715.5E6	690.391	641.726
10) T	Pentachlo...	8.484	9.374	26208.1E6	11519.4E6	720.574	696.779
11) T	2,4,5-TP ...	9.051	9.749	10414.4E6	4132.8E6	711.608	689.278
12) T	2,4,5-T	9.336	10.166	10620.2E6	3413.0E6	709.536	652.974
13) T	2,4-DB	9.900	10.728	1616.5E6	400.0E6	696.099	650.019
14) T	DINOSEB	11.079	11.103	6947.3E6	2760.4E6	710.298	656.375
15) T	Picloram	10.895	12.185	12328.9E6	3300.3E6	659.028	710.157
16) T	DCPA	11.378	12.137	12115.8E6	4229.0E6	729.981	656.157

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091624\
Data File : PS027694.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2024 13:28
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/17/2024
Supervised By : Ankita Jodhani 09/17/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 03:25:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
Quant Title : 8080.M
QLast Update : Sat Sep 14 02:36:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

