

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS091718.M
 Title : 8080.M
 Last Update : Tue Sep 18 03:01:16 2018
 Response Via : Initial Calibration

Calibration Files

750 =PS001770.D 1000 =PS001771.D 1500 =PS001772.D
 500 =PS001769.D 200 =PS001768.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	3.269	3.118	3.031	3.270	3.571	3.252	E5	6.33
2) T	3,5-DICHLOROBENZOI...	3.322	3.189	3.034	3.313	3.599	3.291	E5	6.31
3) T	4-Nitrophenol	1.445	1.375	1.335	1.422	1.462	1.408	E5	3.70
4) S	2,4-DCAA	2.267	2.137	2.064	2.267	2.480	2.243	E5	7.07
5) T	DICAMBA	8.801	8.446	8.152	8.725	9.211	8.667	E5	4.58
6) T	MCPP	6.643	6.425	6.208	6.328	6.258	6.372	E5	2.70
7) T	MCPA	1.066	1.029	0.984	1.056	1.113	1.049	E6	4.55
8) T	DICHLORPROP	2.328	2.203	2.101	2.353	2.618	2.321	E5	8.38
9) T	2,4-D	2.833	2.650	2.636	2.782	2.880	2.756	E5	3.96
10) T	Pentachlorophenol	3.141	2.975	2.770	3.165	3.316	3.073	E6	6.78
11) T	2,4,5-TP (SILVEX)	1.204	1.145	1.099	1.200	1.268	1.183	E6	5.43
12) T	2,4,5-T	1.224	1.155	1.129	1.209	1.252	1.194	E6	4.23
13) T	2,4-DB	1.660	1.570	1.540	1.638	1.735	1.629	E5	4.73
14) T	DINOSEB	8.388	8.094	7.725	8.420	9.054	8.336	E5	5.87
15) T	Picloram	1.797	1.725	1.694	1.746	1.743	1.741	E6	2.16
16) T	DCPA	1.499	1.435	1.352	1.494	1.578	1.472	E6	5.71

Signal #2 Calibration Files

750 =PS001770.D 1000 =PS001771.D 1500 =PS001772.D
 500 =PS001769.D 200 =PS001768.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	3.020	2.910	2.853	2.917	3.070	2.954	E5	2.99
2) T	3,5-DICHLOROBENZOI...	3.190	3.110	3.002	3.176	3.304	3.156	E5	3.52
3) T	4-Nitrophenol	1.292	1.269	1.229	1.226	1.254	1.254	E5	2.21
4) S	2,4-DCAA	2.588	2.488	2.435	2.530	2.660	2.540	E5	3.44
5) T	DICAMBA	9.160	8.964	8.756	8.790	8.924	8.919	E5	1.80
6) T	MCPP	6.265	6.255	6.238	5.807	4.898	5.893	E5	9.98
7) T	MCPA	9.266	9.290	9.243	8.695	6.847	8.668	E5	12.09
8) T	DICHLORPROP	2.373	2.303	2.239	2.343	2.356	2.323	E5	2.31
9) T	2,4-D	3.154	3.084	3.009	3.123	3.185	3.111	E5	2.20
10) T	Pentachlorophenol	3.515	3.372	3.154	3.513	3.582	3.427	E6	4.98
11) T	2,4,5-TP (SILVEX)	1.404	1.363	1.309	1.378	1.401	1.371	E6	2.81
12) T	2,4,5-T	1.145	1.114	1.080	1.101	1.082	1.104	E6	2.44
13) T	2,4-DB	1.719	1.646	1.661	1.600	1.649	1.655	E5	2.59
14) T	DINOSEB	9.634	9.358	9.057	9.521	9.821	9.478	E5	3.06
15) T	Picloram	1.745	1.687	1.604	1.717	1.751	1.701	E6	3.51
16) T	DCPA	1.919	1.891	1.847	1.824	1.742	1.844	E6	3.68

(#) = Out of Range