

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091718\
 Data File : PS001782.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2018 02:45
 Operator : EI\SJ
 Sample : J4919-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 07:13:30 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 18 03:01:16 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.472	10.969	19109522	19725736	85.190	77.655
Target Compounds							
1) T	Dalapon	2.902	2.381	23737239	22793539	73.000	77.157
5) T	DICAMBA	0.000	11.214	0	21051254	N.D.	23.603 #
6) T	MCPD	12.162	11.502	19459847	5006685	30.538	8.497 #
7) T	MCPA	12.440	0.000	4256114	0	4.055	N.D. #
10) T	Pentachlo...	13.746	13.413	5512335	434568	1.794	0.127 #
12) T	2,4,5-T	15.026	14.571	20271879	13207534	16.981	11.960 #
13) T	2,4-DB	15.757	15.222	14127841	8568684	86.752	51.768 #
14) T	DINOSEB	17.207	15.633	4682548	2613144	5.617	2.757 #
15) T	Picloram	17.035	16.851	3305107	4017853	1.898	2.362



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091718\
Data File : PS001782.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2018 02:45
Operator : EI\SJ
Sample : J4919-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 07:13:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
Quant Title : 8080.M
QLast Update : Tue Sep 18 03:01:16 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.902 min
Delta R.T.: -0.047 min
Response: 23737239
Conc: 73.00 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.381 min
Delta R.T.: -0.029 min
Response: 22793539
Conc: 77.16 ng/ml




#4 2,4-DCAA

R.T.: 11.472 min
Delta R.T.: -0.007 min
Response: 19109522
Conc: 85.19 ng/ml



#4 2,4-DCAA


R.T.: 10.969 min
Delta R.T.: -0.007 min
Response: 19725736
Conc: 77.66 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 11.790 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.214 min
Delta R.T.: -0.045 min
Response: 21051254
Conc: 23.60 ng/ml




#6 MCP

R.T.: 12.162 min
Delta R.T.: 0.037 min
Response: 19459847
Conc: 30.54 ug/ml



#6 MCP


R.T.: 11.502 min
Delta R.T.: 0.019 min
Response: 5006685
Conc: 8.50 ug/ml



#7 MCPA


R.T.: 12.440 min
Delta R.T.: 0.065 min
Response: 4256114
Conc: 4.06 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 11.858 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 13.746 min
Delta R.T.: 0.028 min
Response: 5512335
Conc: 1.79 ng/ml



#10 Pentachlorophenol


R.T.: 13.413 min
Delta R.T.: 0.007 min
Response: 434568
Conc: 0.13 ng/ml



#12 2,4,5-T


R.T.: 15.026 min
Delta R.T.: 0.037 min
Response: 20271879
Conc: 16.98 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 14.571 min
Delta R.T.: -0.020 min
Response: 13207534
Conc: 11.96 ng/ml




#13 2,4-DB

R.T.: 15.757 min
Delta R.T.: 0.020 min
Response: 14127841
Conc: 86.75 ng/ml



#13 2,4-DB


R.T.: 15.222 min
Delta R.T.: -0.018 min
Response: 8568684
Conc: 51.77 ng/ml



#14 DINOSEB


R.T.: 17.207 min
Delta R.T.: -0.016 min
Response: 4682548
Conc: 5.62 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 15.633 min
Delta R.T.: -0.036 min
Response: 2613144
Conc: 2.76 ng/ml




#15 Picloram

R.T.: 17.035 min
Delta R.T.: 0.046 min
Response: 3305107
Conc: 1.90 ng/ml



#15 Picloram


R.T.: 16.851 min
Delta R.T.: 0.000 min
Response: 4017853
Conc: 2.36 ng/ml



#16 DCPA

R.T.: 17.598 min
Delta R.T.: 0.033 min
Response: -6054655
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 16.906 min
Delta R.T.: 0.000 min
Response: 7434340
Conc: 4.03 ng/ml