

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091719\
 Data File : PS006777.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Sep 2019 11:30
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 9/19/2019 9:05:01 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 06:52:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 10 06:14:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.500	5.224	586.4E6	314.0E6	754.855	735.431
Target Compounds							
1) T	Dalapon	2.067	1.774	704.2E6	360.4E6	720.444m	705.146
2) T	3,5-DICHL...	4.866	4.526	740.0E6	368.4E6	747.525	680.733
3) T	4-Nitroph...	5.352	4.909	241.8E6	142.9E6	681.240	645.512
5) T	DICAMBA	5.652	5.351	2077.1E6	1100.2E6	720.485	686.215
6) T	MCPD	5.804	5.434	134.5E6	73055524	70.704	69.030
7) T	MCPA	5.937	5.612	209.1E6	110.8E6	73.977	63.164m
8) T	DICHLORPROP	6.282	5.891	514.4E6	266.4E6	723.320	684.426
9) T	2,4-D	6.496	6.162	599.5E6	332.6E6	775.963	697.346
10) T	Pentachlo...	6.758	6.528	9124.4E6	4538.2E6	819.076	753.236
11) T	2,4,5-TP ...	7.292	6.891	3219.5E6	1623.2E6	748.199	713.053
12) T	2,4,5-T	7.565	7.251	3602.7E6	1571.4E6	773.181	728.774
13) T	2,4-DB	8.095	7.737	440.6E6	196.3E6	724.016	713.823
14) T	DINOSEB	9.184	8.047	2804.5E6	1263.2E6	805.715	753.995
15) T	Picloram	9.018	8.985	4391.3E6	2225.2E6	650.973	601.312
16) T	DCPA	9.455	8.942	4361.1E6	2280.2E6	771.995	723.833

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091719\
Data File : PS006777.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Sep 2019 11:30
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
9/19/2019 9:05:01 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 06:52:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090919.M
Quant Title : 8080.M
QLast Update : Tue Sep 10 06:14:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

