

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091720\
 Data File : PS012342.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Sep 2020 16:01
 Operator : DD\AJ
 Sample : L4021-04MS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 B15-5-7MS

Manual Integrations
 APPROVED

Ankita
 9/18/2020 10:09:28 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 05:39:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 15 06:03:04 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.215	6.911	513.8E6	236.0E6	474.012	447.395
Target Compounds							
1) T	Dalapon	3.030	2.623	262.8E6	69258807	214.066	104.045m#
2) T	3,5-DICHL...	6.430	6.064	407.4E6	216.8E6	277.827	268.028m
3) T	4-Nitroph...	7.048	6.523	82423137	46205842	154.775	144.456
5) T	DICAMBA	7.406	7.079	1143.1E6	597.0E6	271.672	276.828
6) T	MCPD	7.571	7.171	97052221	43502842	30.120	25.278
7) T	MCPA	7.720	7.385	138.0E6	105.6E6	29.534	37.563 #
8) T	DICHLORPROP	8.098	7.705	282.1E6	160.9E6	275.886	284.412
9) T	2,4-D	8.328	8.000	393.5E6	235.4E6	345.483	323.542
10) T	Pentachlo...	8.637	8.428	4098.5E6	2033.1E6	278.634	300.193
11) T	2,4,5-TP ...	9.186	8.788	1805.1E6	893.3E6	306.708	308.660
12) T	2,4,5-T	9.472	9.169	1730.8E6	859.2E6	282.210	298.328
13) T	2,4-DB	10.025	9.675	234.2E6	166.8E6	306.107	410.007 #
14) T	DINOSEB	11.182	10.007	365.6E6	173.1E6	77.857	83.358
15) T	Picloram	11.005	10.990	2126.4E6	1250.3E6	288.790	294.911
16) T	DCPA	11.449	10.934	1409.8E6	837.6E6	190.508	230.438

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091720\
Data File : PS012342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Sep 2020 16:01
Operator : DD\AJ
Sample : L4021-04MS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
B15-5-7MS

Manual Integrations
APPROVED

Ankita
9/18/2020 10:09:28 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 05:39:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
Quant Title : 8080.M
QLast Update : Tue Sep 15 06:03:04 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

