

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091923\
 Data File : PS024342.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 09:44
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/20/2023
 Supervised By : Ankita Jodhani 09/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 01:40:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 28 16:48:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.362	6.787	1500.2E6	502.8E6	753.003	748.613
Target Compounds							
1) T	Dalapon	2.483	2.546	1475.8E6	690.6E6	679.012	654.940
2) T	3,5-DICHL...	5.736	5.971	2139.4E6	735.4E6	703.816	705.291
3) T	4-Nitroph...	6.199	6.401	866.3E6	281.9E6	697.295	668.375
5) T	DICAMBA	6.503	6.950	6289.9E6	2073.9E6	721.553	720.079
6) T	MCP	6.655	7.046	436.6E6	139.8E6	70.112	72.990m
7) T	MCPA	6.777	7.253	624.0E6	201.7E6	71.888	68.996
8) T	DICHLORPROP	7.091	7.565	1587.2E6	499.9E6	698.187	687.989
9) T	2,4-D	7.288	7.849	1863.0E6	625.0E6	700.391	701.585
10) T	Pentachlo...	7.518	8.274	23148.3E6	7694.5E6	740.209	726.262
11) T	2,4,5-TP ...	8.039	8.632	9699.5E6	2981.8E6	725.640	718.651
12) T	2,4,5-T	8.293	9.003	8985.3E6	2872.3E6	725.202	715.745
13) T	2,4-DB	8.798	9.506	1519.5E6	356.7E6	704.899	776.153
14) T	DINOSEB	9.853	9.834	6874.6E6	2117.2E6	711.479	708.273
15) T	Picloram	9.678	10.787	11515.1E6	4153.6E6	705.096	699.408
16) T	DCPA	10.118	10.757	11739.7E6	3761.0E6	724.942	733.139

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091923\
Data File : PS024342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2023 09:44
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/20/2023
Supervised By : Ankita Jodhani 09/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 01:40:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082823.M
Quant Title : 8080.M
QLast Update : Mon Aug 28 16:48:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

