

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091924\
 Data File : PS027726.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2024 16:58
 Operator : AR\AJ
 Sample : PB163515BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB163515BS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/20/2024
 Supervised By :Ankita Jodhani 09/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 01:04:02 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 14 02:36:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.109	7.632	1314.9E6	544.4E6	528.574m	549.368m
Target Compounds							
1) T	Dalapon	2.558	2.626	1662.4E6	622.5E6	401.240	381.453
2) T	3,5-DICHL...	6.299	6.603	1802.0E6	785.4E6	493.314	487.383
3) T	4-Nitroph...	6.905	7.167	775.5E6	308.3E6	458.033	430.068
5) T	DICAMBA	7.289	7.827	4735.6E6	1972.3E6	468.629	447.593
6) T	MCPD	7.467	7.928	289.4E6	159.1E6	41.982	48.477
7) T	MCPA	7.612	8.168	464.3E6	306.0E6	45.836	57.812 #
8) T	DICHLORPROP	7.976	8.533	1318.4E6	585.0E6	497.329	523.636
9) T	2,4-D	8.199	8.858	1619.5E6	660.5E6	528.333	592.388
10) T	Pentachlo...	8.486	9.374	18582.5E6	7843.1E6	510.914	474.412
11) T	2,4,5-TP ...	9.052	9.749	7450.1E6	2997.2E6	509.061	499.876
12) T	2,4,5-T	9.338	10.166	7258.1E6	2442.4E6	484.913	467.285
13) T	2,4-DB	9.902	10.727	1204.0E6	311.7E6	518.470	506.535
14) T	DINOSEB	11.080	11.104	4444.7E6	1884.7E6	454.430	448.153
15) T	Picloram	10.896	12.182	7425.2E6	1963.5E6	396.905	422.490
16) T	DCPA	11.378	12.136	8200.4E6	3404.2E6	494.078	528.186

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091924\
Data File : PS027726.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2024 16:58
Operator : AR\AJ
Sample : PB163515BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB163515BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/20/2024
Supervised By : Ankita Jodhani 09/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 01:04:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
Quant Title : 8080.M
QLast Update : Sat Sep 14 02:36:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

