

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092019\
 Data File : PS006838.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2019 10:34
 Operator : SM\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 11:04:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 10:42:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.518	5.249	864.0E6	438.9E6	899.350	922.197
Target Compounds							
1) T	Dalapon	2.090	1.799	967.3E6	463.1E6	902.584	875.198
2) T	3,5-DICHL...	4.886	4.552	1003.8E6	523.0E6	903.806	890.470
3) T	4-Nitroph...	5.371	4.935	355.6E6	210.5E6	950.246	859.190
5) T	DICAMBA	5.670	5.377	2939.3E6	1551.4E6	881.172	902.289
6) T	MCPD	5.823	5.462	208.9E6	111.1E6	102.126	97.401
7) T	MCPA	5.956	5.640	295.9E6	165.0E6	97.016	87.769
8) T	DICHLORPROP	6.298	5.916	705.1E6	378.0E6	898.485	891.404
9) T	2,4-D	6.514	6.187	776.1E6	468.2E6	946.703	898.924
10) T	Pentachlo...	6.774	6.553	10879.0E6	5778.5E6	867.871	877.095
11) T	2,4,5-TP ...	7.309	6.916	4192.2E6	2213.3E6	893.317	898.571
12) T	2,4,5-T	7.581	7.275	4658.4E6	2074.5E6	874.093	965.711
13) T	2,4-DB	8.112	7.761	555.4E6	281.9E6	973.513	1022.331
14) T	DINOSEB	9.200	8.071	3359.3E6	1605.5E6	838.646	920.240
15) T	Picloram	9.033	9.009	6516.9E6	3488.9E6	985.965	966.492
16) T	DCPA	9.472	8.967	5344.3E6	2958.9E6	874.798	905.765

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092019\
Data File : PS006838.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2019 10:34
Operator : SM\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 11:04:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 10:42:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

