

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092019\
 Data File : PS006867.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2019 21:14
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 23:57:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 11:14:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.500	5.231	686.3E6	342.2E6	751.038	744.098
Target Compounds							
1) T	Dalapon	2.071	1.781	729.8E6	360.2E6	683.055	675.724
2) T	3,5-DICHL...	4.868	4.535	771.3E6	404.7E6	709.068	703.919
3) T	4-Nitroph...	5.354	4.917	272.9E6	162.8E6	715.725	673.397
5) T	DICAMBA	5.652	5.359	2318.2E6	1210.5E6	717.485	719.130
6) T	MCPD	5.804	5.442	159.4E6	85386098	74.913	73.512
7) T	MCPA	5.937	5.620	229.2E6	130.0E6	73.671	70.604
8) T	DICHLORPROP	6.280	5.898	552.3E6	295.7E6	719.283	714.137
9) T	2,4-D	6.496	6.169	611.6E6	367.2E6	746.296	719.921
10) T	Pentachlo...	6.756	6.535	8773.0E6	4591.1E6	736.106	729.733
11) T	2,4,5-TP ...	7.290	6.898	3361.3E6	1754.1E6	739.831	732.277
12) T	2,4,5-T	7.563	7.257	3706.9E6	1624.1E6	728.517	751.393
13) T	2,4-DB	8.094	7.744	449.8E6	207.9E6	775.587	730.762
14) T	DINOSEB	9.181	8.053	2702.8E6	1263.3E6	711.644	733.293
15) T	Picloram	9.016	8.991	4831.8E6	2515.0E6	720.555	691.003
16) T	DCPA	9.454	8.948	4350.3E6	2358.7E6	745.505	742.315

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092019\
Data File : PS006867.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2019 21:14
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 23:57:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 11:14:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

