

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092021\
 Data File : PS016845.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2021 19:39
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS092021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 01:53:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 21 01:52:13 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	2.384	2.502 E6	-4.9	103	0.00
2 T	3,5-DICHLORO BENZOIC ACID	2.128	2.114 E6	0.7	98	0.00
3 T	4-Nitrophenol	866.263	843.768 E3	2.6	95	0.00
4 S	2,4-DCAA	1.524	1.515 E6	0.6	98	0.00
5 T	DICAMBA	5.836	5.689 E6	2.5	95	0.00
6 T	MCP P	4.333	4.281 E6	1.2	94	0.00
7 T	MCPA	6.185	6.216 E6	-0.5	98	0.00
8 T	DICHLORPROP	1.476	1.435 E6	2.8	96	0.00
9 T	2,4-D	1.758	1.772 E6	-0.8	99	0.00
10 T	Pentachlorophenol	20.206	20.329 E6	-0.6	99	0.00
11 T	2,4,5-TP (SILVEX)	7.985	8.087 E6	-1.3	99	0.00
12 T	2,4,5-T	8.254	8.343 E6	-1.1	99	0.00
13 T	2,4-DB	1.653	1.670 E6	-1.0	98	0.00
14 T	DINOSEB	6.093	6.152 E6	-1.0	99	0.00
15 T	Picloram	11.902	12.182 E6	-2.4	99	0.00
16 T	DCPA	10.268	9.379 E6	8.7	90	0.00

Signal #2

1 T	Dalapon	726.829	761.672 E3	-4.8	103	0.00
2 T	3,5-DICHLORO BENZOIC ACID	881.636	897.175 E3	-1.8	98	0.00
3 T	4-Nitrophenol	312.725	310.376 E3	0.8	97	0.00
4 S	2,4-DCAA	586.363	587.892 E3	-0.3	98	0.00
5 T	DICAMBA	2.547	2.525 E6	0.9	96	0.00
6 T	MCP P	2.049	2.001 E6	2.3	94	0.00
7 T	MCPA	3.045	3.019 E6	0.9	98	0.00
8 T	DICHLORPROP	648.523	638.502 E3	1.5	96	0.00
9 T	2,4-D	760.834	774.588 E3	-1.8	99	0.00
10 T	Pentachlorophenol	8.229	8.367 E6	-1.7	99	0.00
11 T	2,4,5-TP (SILVEX)	3.501	3.597 E6	-2.7	100	0.00
12 T	2,4,5-T	3.254	3.358 E6	-3.2	99	0.00
13 T	2,4-DB	416.674	419.715 E3	-0.7	98	0.00
14 T	DINOSEB	2.450	2.550 E6	-4.1	100	0.00
15 T	Picloram	4.290	4.453 E6	-3.8	99	0.00
16 T	DCPA	4.359	4.019 E6	7.8	90	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092021\
Data File : PS016845.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2021 19:39
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS092021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 01:53:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092021.M
Quant Title : 8080.M
QLast Update : Tue Sep 21 01:52:13 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0