

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092120\
 Data File : PS012396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2020 09:34
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 10:39:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 15 06:03:04 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
2) T	3,5-DICHL...	6.475	0.000	36990405	0	25.229	N.D. #
12) T	2,4,5-T	9.420	0.000	9335003	0	1.522	N.D. #
15) T	Picloram	10.978	11.034	46439849	88713696	6.307	20.926 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091210\
Data File : PS012396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2020 09:34
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 10:39:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
Quant Title : 8080.M
QLast Update : Tue Sep 15 06:03:04 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.475 min
Delta R.T.: 0.044 min
Response: 36990405
Conc: 25.23 ng/ml

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.065 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 7.216 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 6.912 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 9.420 min
Delta R.T.: -0.053 min
Response: 9335003
Conc: 1.52 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T. : 9.170 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 10.978 min
Delta R.T.: -0.031 min
Response: 46439849
Conc: 6.31 ng/ml



#15 Picloram

R.T.: 11.034 min
Delta R.T.: 0.039 min
Response: 88713696
Conc: 20.93 ng/ml