

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092120\
 Data File : PS012414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2020 21:37
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 9/22/2020 9:32:34 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 02:44:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 15 06:03:04 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.214	6.911	829.1E6	382.6E6	764.910	725.242m
Target Compounds							
1) T	Dalapon	3.033	2.623	837.3E6	438.3E6	682.079	658.445
2) T	3,5-DICHL...	6.429	6.064	1033.9E6	546.1E6	705.134	675.053
3) T	4-Nitroph...	7.046	6.521	368.3E6	213.7E6	691.576	668.098
5) T	DICAMBA	7.404	7.079	3043.6E6	1503.6E6	723.347	697.136
6) T	MCPD	7.573	7.176	237.3E6	120.3E6	73.638	69.921
7) T	MCPA	7.723	7.391	331.4E6	184.9E6	70.930	65.777
8) T	DICHLORPROP	8.096	7.705	734.4E6	389.4E6	718.082	688.253
9) T	2,4-D	8.326	8.000	819.0E6	499.1E6	719.061	686.144
10) T	Pentachlo...	8.635	8.427	10691.0E6	4765.0E6	726.818	703.557
11) T	2,4,5-TP ...	9.184	8.788	4304.3E6	2047.0E6	731.338	707.307
12) T	2,4,5-T	9.470	9.168	4435.1E6	2061.8E6	723.147	715.866
13) T	2,4-DB	10.023	9.675	547.0E6	286.9E6	714.803	705.247
14) T	DINOSEB	11.180	10.006	3373.2E6	1502.8E6	718.298	723.584
15) T	Picloram	11.007	10.995	3863.4E6	2165.4E6	524.709	510.778
16) T	DCPA	11.445	10.933	5397.9E6	2606.7E6	729.410	717.168

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS091210\
Data File : PS012414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2020 21:37
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
9/22/2020 9:32:34 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 02:44:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091420.M
Quant Title : 8080.M
QLast Update : Tue Sep 15 06:03:04 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

