

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092122\
 Data File : PS020630.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2022 17:50
 Operator : AJ\MA
 Sample : N4702-01MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BU-03-091622MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/22/2022
 Supervised By : Ankita Jodhani 09/22/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 23:59:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 14 02:51:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.172	6.757	160.2E6	56694972	46.557m	66.567 #
Target Compounds							
1) T	Dalapon	2.999	2.514	902.3E6	315.3E6	179.765	260.030m#
2) T	3,5-DICHL...	6.395	5.948	159.5E6	27640788	31.991m	20.255m#
3) T	4-Nitroph...	7.013	6.395	47536171	21849095	151.820	46.368 #
5) T	DICAMBA	7.361	6.918	951.1E6	185.3E6	69.551m	48.124m#
6) T	MCPD	7.526	7.011	26109482	26420406	2.507	9.149 #
7) T	MCPA	7.676	7.217	76932840	34557916	5.187	8.004 #
8) T	DICHLORPROP	8.055	7.540	164.7E6	69148491	45.680	68.534 #
9) T	2,4-D	8.288	7.832	302.7E6	82627647	65.115	65.059m
10) T	Pentachlo...	8.588	8.252	746.5E6	174.9E6	16.185	12.345
11) T	2,4,5-TP ...	9.137	8.613	462.6E6	115.1E6	22.631	19.446
12) T	2,4,5-T	9.422	8.990	604.7E6	161.8E6	27.304	29.685
15) T	Picloram	10.955	10.801	5989.3E6	1407.4E6	199.024	174.767
16) T	DCPA	11.393	10.744	6067.7E6	2179.1E6	236.850	295.451

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092122\
Data File : PS020630.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2022 17:50
Operator : AJ\MA
Sample : N4702-01MS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BU-03-091622MS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/22/2022
Supervised By : Ankita Jodhani 09/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 23:59:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091322.M
Quant Title : 8080.M
QLast Update : Wed Sep 14 02:51:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

