

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092220\
 Data File : PS012432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2020 16:24
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 05:05:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	1.136	1.141 E6	-0.4	103	0.00
2 T	3,5-DICHLORO BENZOIC ACID	1.348	1.349 E6	-0.1	103	0.00
3 T	4-Nitrophenol	486.264	492.126 E3	-1.2	104	0.00
4 S	2,4-DCAA	979.449	981.630 E3	-0.2	104	0.00
5 T	DICAMBA	3.858	3.888 E6	-0.8	103	0.00
6 T	MCP	2.918	2.987 E6	-2.4	104	0.00
7 T	MCPA	4.171	4.245 E6	-1.8	104	0.00
8 T	DICHLORPROP	967.692	967.206 E3	0.1	104	0.00
9 T	2,4-D	1.135	1.137 E6	-0.2	104	0.00
10 T	Pentachlorophenol	13.820	13.867 E6	-0.3	103	0.00
11 T	2,4,5-TP (SILVEX)	5.545	5.576 E6	-0.6	103	0.00
12 T	2,4,5-T	5.621	5.658 E6	-0.7	104	0.00
13 T	2,4-DB	946.301	971.966 E3	-2.7	104	0.00
14 T	DINOSEB	4.540	4.509 E6	0.7	103	0.00
15 T	Picloram	7.200	7.417 E6	-3.0	104	0.00
16 T	DCPA	7.063	7.065 E6	-0.0	103	0.00

Signal #2

1 T	Dalapon	598.190	597.884 E3	0.1	102	0.00
2 T	3,5-DICHLORO BENZOIC ACID	712.960	710.115 E3	0.4	103	0.00
3 T	4-Nitrophenol	279.409	283.206 E3	-1.4	103	0.00
4 S	2,4-DCAA	460.017	457.407 E3	0.6	103	0.00
5 T	DICAMBA	1.907	1.916 E6	-0.5	103	0.00
6 T	MCP	1.513	1.528 E6	-1.0	104	0.00
7 T	MCPA	2.397	2.354 E6	1.8	103	0.00
8 T	DICHLORPROP	516.327	511.921 E3	0.9	104	0.00
9 T	2,4-D	626.241	624.560 E3	0.3	103	0.00
10 T	Pentachlorophenol	6.099	6.033 E6	1.1	102	0.00
11 T	2,4,5-TP (SILVEX)	2.638	2.638 E6	0.0	103	0.00
12 T	2,4,5-T	2.571	2.588 E6	-0.7	103	0.00
13 T	2,4-DB	374.369	363.840 E3	2.8	101	0.00
14 T	DINOSEB	1.915	1.897 E6	0.9	102	0.00
15 T	Picloram	3.941	4.017 E6	-1.9	104	0.00
16 T	DCPA	3.325	3.310 E6	0.5	103	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092220\
Data File : PS012432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2020 16:24
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 05:05:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0