

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092320\
 Data File : PS012459.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2020 16:22
 Operator : DD\AJ
 Sample : PB131826BSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB131826BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 17:46:53 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
 Quant Title : 8080.M
 QLast Update : Wed Sep 23 04:57:25 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.913	552.4E6	275.2E6	564.035	598.234
Target Compounds							
1) T	Dalapon	3.029	2.619	568.2E6	462.4E6	499.975	772.978 #
2) T	3,5-DICHL...	6.430	6.065	696.3E6	340.4E6	516.340	477.503
3) T	4-Nitroph...	7.048	6.523	218.6E6	113.0E6	449.504	404.365
5) T	DICAMBA	7.406	7.081	1873.0E6	949.3E6	485.515	497.816
6) T	MCPD	7.572	7.174	162.5E6	76599315	55.691	50.639
7) T	MCPA	7.721	7.388	227.2E6	158.4E6	54.477	66.056
8) T	DICHLORPROP	8.097	7.706	496.4E6	276.5E6	512.935	535.571
9) T	2,4-D	8.328	8.002	615.0E6	378.8E6	541.626	604.958
10) T	Pentachlo...	8.637	8.429	7519.1E6	3431.0E6	544.078	562.516
11) T	2,4,5-TP ...	9.185	8.789	2989.8E6	1472.4E6	539.212	558.070
12) T	2,4,5-T	9.472	9.170	2857.8E6	1385.6E6	508.389	538.866
13) T	2,4-DB	10.024	9.676	332.0E6	197.2E6	350.839	526.692 #
14) T	DINOSEB	11.181	10.007	2009.4E6	923.8E6	442.569	482.269
15) T	Picloram	11.008	10.995	3202.4E6	1855.5E6	444.774	470.766
16) T	DCPA	11.448	10.934	2359.5E6	1319.8E6	334.058	396.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092320\
Data File : PS012459.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2020 16:22
Operator : DD\AJ
Sample : PB131826BSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB131826BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 17:46:53 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092220.M
Quant Title : 8080.M
QLast Update : Wed Sep 23 04:57:25 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

