

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092418\
 Data File : PS001823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2018 03:24
 Operator : EI\SJ
 Sample : J5021-20
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 06:33:41 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 18 03:01:16 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.453	10.949	67310749	65911163	300.072	259.476
Target Compounds							
1) T	Dalapon	0.000	2.364	0	24961130	N.D.	84.495 #
2) T	3,5-DICHL...	0.000	9.105	0	1111528	N.D.	3.522 #
6) T	MCP	12.139	0.000	25823928	0	40.525	N.D. #
15) T	Picloram	0.000	16.881	0	10053274	N.D.	5.911 #
16) T	DCPA	0.000	16.881	0	10053274	N.D.	5.451 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092418\
Data File : PS001823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2018 03:24
Operator : EI\SJ
Sample : J5021-20
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 06:33:41 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091718.M
Quant Title : 8080.M
QLast Update : Tue Sep 18 03:01:16 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 2.949 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.364 min
Delta R.T.: -0.045 min
Response: 24961130
Conc: 84.49 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 9.902 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.105 min
Delta R.T.: -0.021 min
Response: 1111528
Conc: 3.52 ng/ml



#4 2,4-DCAA


R.T.: 11.453 min
Delta R.T.: -0.027 min
Response: 67310749
Conc: 300.07 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 10.949 min
Delta R.T.: -0.027 min
Response: 65911163
Conc: 259.48 ng/ml




#6 MCPP

R.T.: 12.139 min
Delta R.T.: 0.015 min
Response: 25823928
Conc: 40.52 ug/ml



#6 MCPP


R.T.: 0.000 min
Exp R.T.: 11.483 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 0.000 min
Exp R.T. : 16.988 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 16.881 min
Delta R.T.: 0.031 min
Response: 10053274
Conc: 5.91 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 17.565 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 16.881 min
Delta R.T.: -0.025 min
Response: 10053274
Conc: 5.45 ng/ml