

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092519\
 Data File : PS006896.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2019 13:41
 Operator : SM\AJ
 Sample : K5008-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SB-1MS

Manual Integrations
 APPROVED

Ankita
 9/26/2019 5:03:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 01:48:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
 Quant Title : 8080.M
 QLast Update : Fri Sep 20 11:14:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.514	5.247	531.8E6	249.4E6	581.934	542.291
Target Compounds							
1) T	Dalapon	2.082	1.793	360.7E6	211.9E6	337.573m	397.565m
2) T	3,5-DICHL...	4.892	4.550	1356.7E6	187.5E6	1247.284m	326.109m#
3) T	4-Nitroph...	5.370	4.934	67897386	33537577	178.069m	138.740m
5) T	DICAMBA	5.667	5.377	1220.6E6	911.0E6	377.761	541.188m#
6) T	MCPD	5.816	5.458	101.4E6	72554763	47.681m	62.465 #
7) T	MCPA	5.950	5.633	167.7E6	126.8E6	53.913	68.864 #
8) T	DICHLORPROP	6.295	5.914	238.5E6	160.7E6	310.669m	388.094
9) T	2,4-D	6.521	6.184	2261.1E6	271.4E6	2759.059m	532.049m#
10) T	Pentachlo...	6.772	6.551	4097.1E6	2055.2E6	343.773	326.672
11) T	2,4,5-TP ...	7.305	6.913	1537.6E6	770.9E6	338.440	321.826
12) T	2,4,5-T	7.577	7.274	1721.8E6	724.9E6	338.382	335.390
13) T	2,4-DB	8.113	7.760	974.8E6	98429809	1680.575	345.968m#
14) T	DINOSEB	9.197	8.075	497.1E6	336.7E6	130.887	195.415 #
15) T	Picloram	9.032	9.007	2571.4E6	1136.6E6	383.461	312.282
16) T	DCPA	9.469	8.963	1433.6E6	599.9E6	245.675	188.798

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092519\
Data File : PS006896.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2019 13:41
Operator : SM\AJ
Sample : K5008-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
SB-1MS

Manual Integrations
APPROVED

Ankita
9/26/2019 5:03:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 01:48:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS092019.M
Quant Title : 8080.M
QLast Update : Fri Sep 20 11:14:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

