

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092524\
 Data File : PS027767.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2024 13:13
 Operator : AR\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 01:07:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 14 02:36:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.107	7.630	4284749	10414807	1.722	10.509 #
Target Compounds							
1) T	Dalapon	2.529	0.000	24861627	0	6.001	N.D. #
2) T	3,5-DICHL...	6.320	6.580	-456036	520785	N.D.	<MDL
3) T	4-Nitroph...	6.892	7.198f	24137400	2476514	14.256	3.455 #
5) T	DICAMBA	0.000	7.831	0	8112014	N.D.	1.841
6) T	MCPD	7.479	7.926	1101337	5143468	<MDL	1.567 #
7) T	MCPA	7.637	0.000	10088960	0	<MDL	N.D. #
8) T	DICHLORPROP	0.000	8.571f	0	1604924	N.D.	1.437
9) T	2,4-D	8.183	0.000	15750093	0	5.138	N.D. #
10) T	Pentachlo...	8.477	0.000	801871	0	<MDL	N.D. #
11) T	2,4,5-TP ...	9.053	9.750	5987899	1092533	<MDL	<MDL #
12) T	2,4,5-T	9.339	0.000	3249149	0	<MDL	N.D. #
13) T	2,4-DB	9.902	0.000	797163	0	<MDL	N.D. #
14) T	DINOSEB	11.074	11.122	2722177	-44937	<MDL	N.D. #
15) T	Picloram	10.902	12.180	16367847	3642792	<MDL	<MDL #
16) T	DCPA	11.375	12.114f	14003594	5095372	<MDL	<MDL #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092524\
Data File : PS027767.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2024 13:13
Operator : AR\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 01:07:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
Quant Title : 8080.M
QLast Update : Sat Sep 14 02:36:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.529 min
Delta R.T.: -0.025 min
Response: 24861627
Conc: 6.00 ng/ml

Instrument :
ECD_S
ClientSampleId :
HEXANE




#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 2.620 min
Response: 0
Conc: N.D.




#3 4-Nitrophenol

R.T.: 6.892 min
Delta R.T.: -0.014 min
Response: 24137400
Conc: 14.26 ng/ml



#3 4-Nitrophenol


R.T.: 7.198 min
Delta R.T.: 0.030 min
Response: 2476514
Conc: 3.46 ng/ml



#4 2,4-DCAA


R.T.: 7.107 min
Delta R.T.: -0.005 min
Response: 4284749
Conc: 1.72 ng/ml

Instrument :
ECD_S
ClientSampleId :
HEXANE




#4 2,4-DCAA

R.T.: 7.630 min
Delta R.T.: -0.003 min
Response: 10414807
Conc: 10.51 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.291 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 7.831 min
Delta R.T.: 0.003 min
Response: 8112014
Conc: 1.84 ng/ml



#6 MCP


R.T.: 7.479 min
Delta R.T.: 0.008 min
Response: 1101337
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
HEXANE




#6 MCP

R.T.: 7.926 min
Delta R.T.: -0.005 min
Response: 5143468
Conc: 1.57 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 7.978 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 8.571 min
Delta R.T.: 0.036 min
Response: 1604924
Conc: 1.44 ng/ml



#9 2,4-D

R.T.: 8.183 min
Delta R.T.: -0.018 min
Response: 15750093
Conc: 5.14 ng/ml

Instrument :
ECD_S
ClientSampleId :
HEXANE



#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.861 min
Response: 0
Conc: N.D.