

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092524\
 Data File : PS027769.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2024 14:01
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/26/2024
 Supervised By : Ankita Jodhani 09/27/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 02:42:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
 Quant Title : 8080.M
 QLast Update : Sat Sep 14 02:36:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.105	7.627	1894.6E6	717.8E6	761.608	724.323m
Target Compounds							
1) T	Dalapon	2.550	2.618	2607.7E6	1080.1E6	629.403	661.858
2) T	3,5-DICHL...	6.294	6.598	2605.8E6	1108.5E6	713.354m	687.918
3) T	4-Nitroph...	6.899	7.162	1179.1E6	475.9E6	696.433	663.882
5) T	DICAMBA	7.283	7.822	7411.4E6	3046.7E6	733.429	691.433
6) T	MCPD	7.463	7.924	479.9E6	234.9E6	69.621	71.566
7) T	MCPA	7.608	8.164	728.2E6	355.7E6	71.883	67.187
8) T	DICHLORPROP	7.970	8.528	1910.2E6	768.3E6	720.585	687.684
9) T	2,4-D	8.193	8.853	2159.1E6	735.0E6	704.390	659.219
10) T	Pentachlo...	8.479	9.367	27118.8E6	11430.3E6	745.613	691.392
11) T	2,4,5-TP ...	9.046	9.743	10998.5E6	4207.7E6	751.520	701.761
12) T	2,4,5-T	9.331	10.159	10726.0E6	3442.5E6	716.606	658.626
13) T	2,4-DB	9.895	10.720	1580.8E6	387.6E6	680.741	629.861
14) T	DINOSEB	11.072	11.095	7198.6E6	2804.2E6	735.988	666.791
15) T	Picloram	10.889	12.174	10846.8E6	2390.0E6	579.801	514.283
16) T	DCPA	11.371	12.128	12457.7E6	4473.3E6	750.581	694.056

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092524\
Data File : PS027769.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2024 14:01
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/26/2024
Supervised By : Ankita Jodhani 09/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 02:42:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS091224.M
Quant Title : 8080.M
QLast Update : Sat Sep 14 02:36:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

